

# Creation, Characterization, and Assignment of Opsonic Values for a New Pneumococcal OPA Calibration Serum Panel (Ewha QC Sera Panel A) For 13 Serotypes

Robert L. Burton (MS)<sup>1</sup>, Han Wool Kim (MD, PhD)<sup>2,3</sup>, Soyung Lee (MD, PhD)<sup>2,3</sup>, Hun Kim (PhD)<sup>4</sup>, Jee-hyun Seok (MS)<sup>4</sup>, Sang Heon Lee (MS)<sup>5</sup>, Anne Balloch (MS)<sup>6</sup>, Paul Licciardi (BS)<sup>6</sup>, Rachel Marimla (BS)<sup>6</sup>, Sejong Bae (PhD)<sup>1</sup>, Moon H. Nahm (MD)<sup>1</sup>, Kyung-Hyo Kim (MD, PhD)<sup>\*2,3</sup>

<sup>1</sup> Department of Medicine, University of Alabama at Birmingham, Birmingham, Alabama, USA; <sup>2</sup> Department of Pediatrics, Ewha Womans University School of Medicine, Seoul, Republic of Korea; <sup>3</sup> Center for Vaccine Evaluation and Study, Medical Research Institute, Ewha Womans University School of Medicine, Seoul, Republic of Korea; <sup>4</sup> Bio R&D, Global BD, Life Science R&D Center, SK Chemicals, Seongnam-Si, Gyeonggi-do, Republic of Korea; <sup>5</sup> Vaccine R&D Center, Life Sciences R&D, Life Sciences Company, LG Chem Ltd. Daejeon, Republic of Korea; <sup>6</sup> Pneumococcal Research Group, Murdoch Childrens Research Institute, Royal Children's Hospital, Victoria, Australia

## Abstract

## Introduction

**Background.** Opsonophagocytic assays (OPAs) are useful for assessing the immunogenicity of pneumococcal vaccines, especially in the elderly. To reduce the variability in OPA results from different laboratories, sera with known OPA values are needed for assay calibration. Although a serum panel was created by the US FDA, those sera are in limited quantities and are not available. Therefore, a new panel (Ewha QC Sera Panel A) was created and an international collaborative study was conducted to determine consensus values for the 13 serotypes in PCV13.

**Methods.** Sera were collected from 20 healthy adults after PPV23 vaccination, lyophilized, and aliquoted into at least 150 vials per serum. Four laboratories tested the sera five times, with reference serum pool 007sp included in each run. For each result, an unadjusted opsonic index (OI) and a normalized OI (based on 007sp performance) were calculated. The consensus values of both the unadjusted OIs and the normalized OIs were estimated using an ANOVA model.

**Results.** The results for one laboratory differed significantly from those of the other laboratories and were therefore excluded from consensus value determination. Using data from the three remaining laboratories, consensus OIs (both unadjusted and normalized) were determined for each serum sample for 13 serotypes.

**Conclusion.** Ewha QC Serum Panel A will be useful for calibrating pneumococcal OPAs. The sera can be obtained by contacting Kyung-Hyo Kim (kaykim@ewha.ac.kr) or Si Hyung Yoo (yoosh1130@korea.kr) in Korean MFDS.

Conjugate vaccines targeting the pneumococcal capsule have been quite effective in reducing the burden of disease caused by serotypes covered by the current vaccines. However, as the distribution of pneumococcal serotypes changes, new formulations of conjugate vaccines must be developed. Due to the costs and ethical concerns associated with true vaccine efficacy studies, prospective vaccines may be licensed based on immunogenicity. Opsonophagocytic killing assays (OPAs) are often used to assess the relative immunogenicity of a prospective vaccine, especially in adult populations.

Over the past 15 years, significant improvements have been made to pneumococcal OPAs, resulting in reliable assays that are practical for use in support of vaccine trials. However, a 2011 study [1] found that although OPA results from different laboratories correlated reasonably well, the absolute agreement was quite low. In a follow-up study [2], it was shown that normalization with a reference sera (007sp) significantly reduced the inter-laboratory variability.

The follow-up study [2] also yielded a serum panel that could be used to calibrate pneumococcal OPAs. However, the availability of these sera was significantly limited due to low volumes. Thus, an additional international collaborative study involving 4 laboratories familiar with OPAs was devised to: 1) determine consensus values for an OPA calibration panel that could be readily distributed; and 2) determine if normalization of results with 007sp had any impact on the inter-laboratory agreement of OPA results.

## Methods

**Participating laboratories.** For laboratories participated in this study: Ewha Womans University (Seoul, Republic of Korea), Murdoch Childrens Research Institute (Victoria, Australia), SK Chemicals (Gyeonggi-do, Republic of Korea), and the University of Alabama at Birmingham (Birmingham, AL, USA). Labs are listed alphabetically and this order does not correspond to lab designations A through D.

**Sera.** Pneumococcal reference serum pool 007sp has been described previously (1). For the 20 Ewha QC sera, donors were vaccinated with PPV23 and blood was collected on 2 or 3 occasions, with the first collection ~1 month (15 donors) or ~4 years (5 donors) after vaccination.

**Study design.** Each laboratory tested the twenty calibration sera in five separate runs, with 007sp included once in each run.

**OPAs.** All participating laboratories used a multiplexed OPA (MOPA) format. Each laboratory converted raw colony counts to opsonic indexes (OIs) using OpsoTiter template, with an OI defined as the estimated dilution of serum that kills

50% of the target bacteria. Results that failed to meet a laboratory's criteria were indicated as "IR". Samples with undetectable OIs (ie, OIs <8) were assigned a value of "4" and were not normalized.


**Statistical analyses.** Normalized OIs for the calibration sera were obtained using the formula:

$$\text{Normalized OI} = \text{Unadjusted OI} \times \frac{\text{007sp assigned OI}}{\text{007sp OI from run}}$$

To estimate the non-normalized and normalized consensus OIs for the calibration sera, the log transformed OIs (unadjusted or normalized) were fit by serotype and sample using a mixed effects ANOVA model consisting of the random terms Lab and Run(Lab). Consensus OIs and corresponding 95% confidence intervals (CIs) for the calibration sera were obtained by back-transforming the obtained model intercept and its corresponding 95% CI.

## Results

**Figure 1. 007sp Results of Individual Laboratories.** For the 13 target serotypes (x axis), the laboratory-specific geometric mean 007sp OI (GMOI, y axis) is shown for each participating laboratory (see legend). The assigned 007sp OI (solid horizontal lines, [1]), as well as 3-fold deviations (dashed vertical lines) from the assigned value, for each serotype are also shown.




**Table 1. Model-based assessment of the effect of normalization (with Lab D data).** The overall reduction in variability due to normalization is shown for each serotype. Estimates of CVs of the unadjusted and normalized results from the ANOVA are also shown.

	%CV Unadjusted	%CV Normalized	% Variability Reduction
Pn 1	74%	66%	3%
Pn 3	233%	79%	50%
Pn 4	1048%	180%	51%
Pn 5	91%	64%	22%
Pn 6A	316%	130%	33%
Pn 6B	218%	103%	37%
Pn 7F	444%	109%	48%
Pn 9V	514%	176%	36%
Pn 14	348%	162%	31%
Pn 18C	308%	124%	37%
Pn 19A	259%	98%	44%
Pn 19F	451%	123%	45%
Pn 23F	356%	155%	33%

**Table 2. Normalized Calibration Sera Consensus OIs (without Lab D).** For each serum in the panel, the consensus OI and the corresponding 95% confidence interval are shown for the indicated serotypes. Results in red text indicate at least one laboratory reported an irregular result for at least one run.

	Pn 1	Pn 3	Pn 4	Pn 5	Pn 6A	Pn 6B	Pn 7F	Pn 9V	Pn 14	Pn 18C	Pn 19A	Pn 19F	Pn 23F
QC-01 Consensus OI (95% CI)	566 (217, 1475)	463 (329, 652)	108 (80, 145)	213 (133, 342)	2883 (1663, 4989)	2300 (1052, 5026)	1827 (1227, 2719)	5029 (618, 1674)	834 (3250, 7783)	1170 (533, 1304)	555 (624, 2194)	1326 (348, 884)	1326 (931, 1890)
QC-02 Consensus OI (95% CI)	236 (122, 455)	52 (34, 80)	68 (82, 9041)	859 (34, 137)	57 (0, 14744)	5 (1, 30)	2458 (1754, 3445)	388 (23, 6430)	3855 (216, 6830)	330 (189, 575)	3720 (2911, 4754)	879 (522, 1481)	4 (NA)
QC-03 Consensus OI (95% CI)	22 (5, 93)	48 (28, 82)	9 (2, 44)	6 (1, 50)	933 (532, 1634)	1064 (670, 1689)	844 (418, 1704)	1136 (442, 2916)	471 (275, 807)	1126 (454, 2792)	1428 (642, 3174)	208 (156, 279)	1154 (557, 2390)
QC-04 Consensus OI (95% CI)	351 (188, 653)	4 (1, 13)	1276 (1015, 1604)	584 (262, 1300)	4 (NA)	246 (27, 2265)	960 (588, 1566)	42 (2, 869)	1037 (539, 1996)	388 (190, 791)	399 (206, 773)	312 (190, 511)	4 (NA)
QC-05 Consensus OI (95% CI)	16 (4, 63)	3 (1, 12)	72 (1, 4776)	14 (4, 42)	1451 (822, 2561)	1242 (877, 1758)	604 (414, 1747)	850 (380, 959)	316 (128, 777)	141 (43, 456)	382 (204, 717)	226 (142, 359)	412 (224, 759)
QC-06 Consensus OI (95% CI)	1204 (751, 1930)	105 (53, 211)	1511 (1122, 2033)	121 (58, 251)	1802 (1017, 3192)	3177 (1865, 5412)	2259 (1195, 4268)	2733 (1306, 5719)	3679 (1846, 7330)	908 (647, 1275)	844 (577, 1236)	420 (220, 801)	1285 (886, 1864)
QC-07 Consensus OI (95% CI)	2006 (1289, 3122)	181 (140, 233)	3053 (1991, 4679)	420 (262, 673)	1379 (987, 1929)	2383 (1746, 3254)	6686 (5486, 8150)	7029 (3981, 12411)	8938 (6146, 12998)	163 (20, 1346)	957 (791, 1158)	911 (643, 1291)	951 (621, 1458)
QC-08 Consensus OI (95% CI)	308 (198, 479)	117 (75, 182)	1438 (1119, 1847)	13 (5, 37)	12312 (7428, 20408)	9392 (7122, 12384)	1201 (780, 1850)	4541 (3220, 6404)	1352 (718, 2547)	1057 (493, 2266)	1128 (672, 1893)	636 (316, 1282)	2643 (1701, 4106)
QC-09 Consensus OI (95% CI)	21 (3, 153)	179 (112, 287)	2269 (712, 7232)	52 (19, 141)	677 (452, 1015)	720 (459, 1129)	8644 (5031, 14854)	1834 (773, 4354)	1325 (955, 1837)	276 (0, 3451)	392 (208, 737)	999 (439, 2274)	688 (363, 1303)
QC-10 Consensus OI (95% CI)	1180 (493, 2823)	112 (68, 185)	841 (524, 1351)	30 (8, 111)	902 (449, 1812)	3079 (2011, 4715)	3969 (600, 1455)	934 (2914, 5405)	1486 (868, 2543)	376 (143, 991)	287 (223, 369)	215 (153, 301)	1489 (806, 2752)
QC-11 Consensus OI (95% CI)	1362 (827, 2242)	32 (20, 54)	652 (370, 1151)	56 (18, 177)	503 (216, 1173)	619 (327, 1169)	874 (552, 1382)	3195 (1220, 8369)	1082 (597, 1959)	6355 (2918, 13839)	447 (236, 845)	644 (266, 1557)	644 (266, 1557)
QC-12 Consensus OI (95% CI)	441 (294, 663)	108 (77, 150)	7105 (1694, 29798)	271 (176, 416)	374 (114, 1221)	2254 (1223, 2659)	2534 (1648, 3897)	13646 (8759, 21257)	1372 (629, 2989)	641 (410, 1002)	808 (344, 1901)	334 (141, 738)	334 (141, 738)
QC-13 Consensus OI (95% CI)	907 (498, 1655)	136 (83, 200)	1326 (755, 2329)	580 (241, 1397)	2339 (1299, 4212)	2407 (1745, 3320)	9645 (5145, 18081)	1545 (628, 3799)	8030 (3593, 17949)	1387 (847, 2272)	2358 (1358, 4094)	1394 (853, 2278)	310 (11, 8749)
QC-14 Consensus OI (95% CI)	3752 (227, 607)	61 (30, 125)	318 (319, 1605)	589 (47, 116)	7290 (24, 286)	3981 (584, 1212)	3242 (684, 1956)	3655 (139, 764)	5204 (1213, 2680)	7225 (164, 528)	2056 (77, 539)	1730 (163, 519)	1538 (108, 459)
QC-15 Consensus OI (95% CI)	565 (309, 1034)	201 (97, 419)	3729 (2643, 5260)	184 (109, 313)	790 (270, 2307)	922 (383, 2221)	1777 (990, 3189)	4817 (2506, 9259)	2509 (929, 6780)	967 (511, 1829)	725 (377, 1394)	496 (291, 847)	258 (14, 4850)
QC-16 Consensus OI (95% CI)	495 (333, 737)	238 (155, 366)	2665 (1859, 3819)	156 (90, 273)	2035 (1274, 3251)	3730 (2691, 5170)	4602 (2327, 9100)	6068 (3649, 10091)	24471 (14373, 41662)	1006 (503, 2014)	3174 (2380, 4233)	1011 (535, 1908)	697 (484, 1004)
QC-17 Consensus OI (95% CI)	756 (468, 1220)	61 (30, 125)	716 (319, 1605)	74 (47, 116)	83 (24, 286)	841 (584, 1212)	1157 (684, 1956)	325 (139, 764)	1803 (1213, 2680)	295 (164, 528)	204 (77, 539)	291 (163, 519)	223 (108, 459)
QC-18 Consensus OI (95% CI)	142 (83, 243)	103 (56, 182)	700 (478, 1025)	15 (8, 31)	6349 (3469, 11620)	5964 (3676, 9677)	14734 (9111, 23829)	3452 (969, 12303)	8147 (3168, 20949)	1970 (1412, 2749)	819 (655, 1023)	575 (388, 851)	1805 (904, 3603)
QC-19 Consensus OI (95% CI)	400 (238, 671)	121 (82, 179)	274 (45, 1693)	248 (203, 305)	443 (98, 2007)	4673 (2829, 7718)	1428 (808, 2524)	922 (179, 4748)	282 (73, 1086)	694 (539, 894)	38 (11, 136)	288 (156, 530)	480 (255, 904)
QC-20 Consensus OI (95% CI)	421 (220, 806)	68 (39, 117)	1291 (802, 2076)	149 (74, 301)	1446 (771, 2713)	3417 (1914, 6100)	4508 (2687, 7565)	3510 (1713, 7191)	6937 (4081, 11792)	631 (418, 952)	817 (434, 1537)	1997 (1436, 2777)	104 (1, 79)

Pn, pneumococcal serotype; OI, Opsonic Index; CI, Confidence Interval; NA, Not Applicable (all reported values were undetectable and/or irregular)



## Conclusions

- The Ewha QC Sera Panel A sera will be an important tool for setting up new OPAs and for calibrating existing OPAs.
- As with a previous study [2], normalization with 007sp decreased the interlaboratory variability between the 4 laboratories. However, the absolute variability remained high for some serotypes, mostly driven by the Lab D data. For many serotypes, the 007sp results from Lab D were quite lower than the assigned values, indicating that perhaps a minimum level of agreement with the assigned 007sp values may be needed to full reap the benefits of normalization.

## References

- Rose CE, et al. Multilaboratory comparison of Streptococcus pneumoniae opsonophagocytic killing assays and their level of agreement for the determination of functional antibody activity in human reference sera. Clin Vaccine Immunol 18:135-142.
- Burton RL, et al. Assignment of Opsonic Values to Pneumococcal Reference Serum 007sp for Use in Opsonophagocytic Assays for 13 Serotypes. Clin Vaccine Immunol 24:01-13.

## Acknowledgements

This work was funded by:  
Ministry of Food and Drug Safety 15172MFDS275 and 17172MFDS275 (K-HK)  
US NIH contract HHSN272201200005C (MHN)